



Store at: +2+8°C.

Presentation:

Cod. SU001 CONT: R 2 x 125 mL.+ CAL 1 x 5 mL.
 Cod. SU001-SP CONT: R 2 x 50 mL.+ CAL 1 x 5 mL.
 Cod. SU001-B CONT: R 8 x 125 mL.+ CAL 1 x 5 mL.

Procedure

Quantitative determination of albumin.

Only for in vitro use in clinical laboratory (IVD)

TEST SUMMARY

Albumin in the presence of bromocresol green at a slightly acid pH, produces a colour change of the indicator from yellow-green to green-blue. The intensity of the colour formed is proportional to the albumin concentration in the sample^{1,2,3,4}.

REAGENTS COMPOSITION

R	Bromocresol green pH 4.2	0.12 mmol/L
Albumin Cal	Albumin aqueous primary calibrator	5 g/dL

REAGENT PREPARATION AND STABILITY

Reagent (R) and standard (Albumin Cal) are ready to use. All the components of the kit are stable until the expiration date on the label when stored at 2-8°C, protected from light and contamination prevented during their use.

Do not use reagents over the expiration date.

Albumin Cal: Proceed carefully with this product because due its nature it can get contaminated easily.

Signs of Reagent deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 630 nm. ≥ 0.40

All the reagents of the kit are stable up to the end of the indicated month and year of expiry. Store tightly closed at 2-8°C. Do not use reagents over the expiration date.

SPECIMEN

Serum or plasma, free of hemolysis¹. Stable 1 month at 2-8°C or 1 week at 15-25°C.

The samples with particles or fibrin should be centrifuged to eliminate them. Do not use haemolized or lipemic samples.

MATERIAL REQUIRED BUT NOT PROVIDED

- Spectrophotometer or colorimeter measuring at 630 nm.
- Matched cuvettes 1.0 cm. light path.

General laboratory equipment.

TEST PROCEDURE

- Assay Conditions
 - Wavelength: 630 nm.
 - Cuvette: 1 cm light path.
 - Temperature 15-25°C.
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Calibrator	Sample
R.1 (mL.)	1.0	1.0	1.0
Calibrator (note 1)(μ L.)	--	5	--
Sample (μ L.)	--	--	5

- Mix and incubate for 10 minutes at room temperature.
- Read the absorbance (A) of the samples and calibrator, against the Blank. The colour is stable 1 hour at room temperature (15-25°C).

CALCULATIONS

$$\text{Albumin in the sample (g/dL.)} = \frac{(A) \text{ Sample} - (A) \text{ Blank}}{(A) \text{ Standard} - (A) \text{ Blank}} \times 5 \text{ (Cal. conc.)}$$

Conversion Factor. g/dL. x 144,9 = μ mol/L.

QUALITY CONTROL

Control sera are recommended to monitor the performance of the procedure, Control H Normal Ref. QC003 and Control H Pathological Ref. QC004. If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

Serum controls are recommended for internal quality control. Each laboratory should establish its own Quality Control scheme and corrective actions.

REFERENCE VALUES

3.5 to 5.0 g/dL¹.

(These values are for orientation purpose).

It is suggested that each laboratory establish its own reference range.

CLINICAL SIGNIFICANCE

One of the most important serum proteins produced in the liver is albumin. This molecule has an extraordinary wide range of functions, including nutrition, maintenance of osmotic pressure and transport of Ca⁺⁺, Bilirubin, free fatty acid, drugs and steroids. Variation in albumin levels indicate liver diseases malnutrition, skin lesions such as dermatitis and burns or dehydration^{1,7,8}. Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

REAGENT PERFORMANCE

Measuring Range:

From detection limit Of 0.0349 g/dL. to linearity limit of 6 g/dL., under the described assay conditions.

If results obtained were greater than linearity limit, dilute the sample $\frac{1}{2}$ with NaCl 9 g/L. and multiply result by 2.

Precision:

Mean (g/dL)	Intra-assay n= 20		Inter-assay n= 20	
	4.17	2.84	4.56	3.07
SD	0.02	0.01	0.28	0.18
CV	0.42	0.53	6.20	5.90

Sensitivity:

1 g/dL. = 0.2003A

Accuracy:

Results obtained GPL reagents did not show systematic differences when compared with other commercial reagents.

The results obtained using 50 samples were the following:

Correlation coefficient (r): 0.99169

Regression Equation : y= 1.045x - 0.028

The results of the performance characteristics depend on the analyzer used.

INTERFERING SUBSTANCES

Interference:

Bilirubin up to 110 mg/L, hemoglobin up to 1 g/L. and lipemic sera up to 10 g/L. no interfere.

Other substances may interfere. A list of drugs and other substances that could interfere has been reported^{5,6}.

NOTES

- Calibration with the aqueous standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.
- Use clean disposable pipette tips for its dispensation.

BIBLIOGRAPHY

- Gendler S,Uric acid. Kaplan A et al. Clin Chem The C.V. Mosby Co. St. Louis Toronto Priceton 1984; 1268-1273 and 425.
- Rodkey F L. Clin Chem. 1965; 11: 478-487.
- Webster D. Clin Chem. 1974; Acta 53: 109-115
- Doumas BT Clin Chem. 1971; Acta 31: 87-96
- Young DS. Effects of drugs on Clinical lab. Tests, 4th ed AACCC Press, 1995.
- Young DS. Effects of disease on Clinical lab. Tests, 4th ed AACCC 2001
- Burtis A et al. Tietz Textbook of Clinical Chemistry, 3rd ed AACCC1999.
- Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed AACCC 1995.

