



Store at: +2+8°C.

Presentation:

Cod. SU027

CONT: R 2 x 125 mL.+ CAL 1 x 5 mL.

Cod. SU027-SP

CONT: R 2 x 50 mL.+ CAL 1 x 5 mL.

Procedure

**Quantitative determination of phosphorus.**

**Only for in vitro use in clinical laboratory (IVD)**

**TEST SUMMARY**

Direct method for determining inorganic phosphate. Inorganic phosphate reacts in acid medium with ammonium molybdate to form a phosphomolybdate complex with yellow colour. The intensity of the colour formed is proportional to the inorganic phosphorus concentration in the sample<sup>1,2</sup>.

**REAGENTS COMPOSITION**

<b>R (Molybdic)</b>	Ammonium molybdate Sulphuric acid (SO <sub>4</sub> H <sub>2</sub> ) Detergents	0.40 mM. 210 mM.
<b>Phosphorus Cal</b>	Phosphorus aqueous primary calibrator	5 mg/dL.

**PRECAUTIONS**

R: H314-Causes severe skin burns and eye damage. Follow the precautionary statements given in MSDS and label of the product.

**REAGENT PREPARATION AND STABILITY**

Reagent (R) and calibrator (Phosphorus Cal) are ready to use. All the components of the kit are stable until the expiration date on the label when stored at 2-8°C, protected from light and contamination prevented during their use.

Do not use reagents over the expiration date.

**Phosphorus Cal:** Proceed carefully with this product because due its nature it can get contaminated easily.

**Signs of Reagent deterioration:**

- Presence of particles and turbidity.
- Blank absorbance (A) at 340 nm. ≥ 0.54

All the reagents of the kit are stable up to the end of the indicated month and year of expiry. Store tightly closed at 2-8°C. Do not use reagents over the expiration date.

**SPECIMEN**

- Serum or plasma<sup>1,5</sup>.
- Free of hemolysis. Serum should be removed from the clot as quickly as possible to avoid elevation of serum phosphorus from hydrolysis or leakage of phosphate present in erythrocytes. Stability: 7 days at 2-8°C.
- Urine<sup>1,2</sup> (24 h):  
Collect the specimen into a bottle containing 10 mL of 10% v/v hydrochloric acid (HCl) to avoid phosphate precipitations. Adjust to pH 2. Dilute the sample 1/10 with distilled water. Mix. Multiply the result by 10 (dilution factor). Stability: 10 days at 2-8°C.

**MATERIAL REQUIRED BUT NOT PROVIDED**

- Spectrophotometer or colorimeter measuring at 340 nm.
- Matched cuvettes 1.0 cm. light path.

General laboratory equipment<sup>(note 1)</sup>.

**TEST PROCEDURE**

- Assay Conditions
  - Wavelength : ..... 340 nm.
  - Cuvette: ..... 1 cm light path.
  - Temperature .....37 / 30 / 25°C.
- Adjust the instrument to zero with distilled water.
- Pipette into a cuvette:

	Blank	Calibrator	Sample
R.1 (mL.)	1.0	1.0	1.0
Calibrator <sup>(Note 2-3)</sup> (µL.)	--	10	--
Sample (µL.)	--	--	10

- Mix and incubate for 5 minutes.
- Read the absorbance (A) of the samples and Standard, against the Blank. The colour is stable for at least 30 minutes.

**CALCULATIONS**

SERUM:

$$\text{Phosphorus (mg/dL.)} = \frac{(A)\text{Sample}}{(A)\text{Standard}} \times 5 \text{ (Standard conc.)}$$

Urine 24 h:

$$\text{Phosphorus (mg/24h.)} = \frac{(A)\text{Sample}}{(A)\text{Standard}} \times 5 \times \text{vol. (dL) urine 24 h}$$

**Conversion Factor.** mg/dL x 0.323 = mmol/L.

**QUALITY CONTROL**

Control sera are recommended to monitor the performance of the procedure, Control H Normal Ref. QC003 and Control H Pathological Ref. QC004. If control values are found outside the defined range, check the instrument, reagents and calibrator for problems.

**Serum controls are recommended for internal quality control. Each laboratory should establish its own Quality Control scheme and corrective actions**

**REFERENCE VALUES**

- Serum or plasma:
  - Children 4.0 – 7.0 mg/dL = 1.29 – 2.26 mmol/L.
  - Adults 2.5 – 5.0 mg/dL = 0.80 – 1.61 mmol/L.
- Urine:
  - Adults 0.4 – 1.3 g /24 h

(These values are for orientation purpose).

It is suggested that each laboratory establish its own reference range.

**CLINICAL SIGNIFICANCE**

Phosphorus is an essential mineral for tissue bone formation and is required by every cell in the body for normal function. Approximately 85% of the body phosphorus is found in bone and in teeth. Low levels of phosphorus, can be caused by hypervitaminosis D, primary hyperparathyroidism, renal tubular disorders, antacids or malabsorption. High levels of phosphorus can be caused by diet, bone metastases, liver disease, alcohol ingestion, diarrhea and vomiting<sup>1,5,6</sup>. Clinical diagnosis should not be made on a single test result; it should integrate clinical and other laboratory data.

**REAGENT PERFORMANCE**

- **Measuring Range:**  
From detection limit of 0.07 mg/dL. to linearity limit of 15 mg/dL., under the described assay conditions. If results obtained were greater than linearity limit, dilute the sample 1/2 with NaCl 9 g/L. and multiply result by 2.
- **Precision:**

Mean (mg/dL)	Intra-assay n= 20		Inter-assay n= 20	
	3.34	3.34	3.45	5.83
SD	0.02	0.04	0.02	0.04
CV	6.64	0.64	0.72	0.68

- **Sensitivity:** 1 mg/dL. = 0.053A
- **Accuracy:**  
Results obtained GPL reagents (y) did not show systematic differences when compared with other commercial reagents (x). The results obtained using 50 samples were the following:  
Correlation coefficient (r): 0.9938  
Regression Equation: y= 0.9902x + 0.0749  
The results of the performance characteristics depend on the analyzer used.

**INTERFERING SUBSTANCES**

- Hemolyzed specimens are unacceptable because erythrocytes contain high concentrations of organic phosphate, which can be hydrolyzed to inorganic phosphate during storage. It increases by 4 to 5 mg/dL per day<sup>5</sup>.
- A list of drugs and other interfering substances has been reported by Young et. al<sup>3,4</sup>.

**NOTES**

- Most of the detergents and water softening products used in the laboratories contain chelating agents and phosphates. It is recommended to rinse glassware in diluted nitric acid and water before using.
- Calibration with the aqueous standard may cause a systematic error in automatic procedures. In these cases, it is recommended to use a serum Calibrator.
- Use clean disposable pipette tips for its dispensation.

**BIBLIOGRAPHY**

- Farrell E C. Phosphorus. Kaplan A et al. Clin Chem The C.V. Mosby Co. St Louis. Toronto. Princeton 1984; 1072-1074 and 418.
- Daly J A. et al. Clin Chem 1972; 18 (3): 263-265.
- Young DS. Effects of drugs on Clin Lab. Tests, 4th ed AACC Press, 1995.
- Young DS. Effects of disease on Clinical Lab. Tests, 4th ed. AACC 2001.
- Burtis A. et al. Tietz Textbook of Clinical Chemistry, 3rd ed. AACC 1999.
- Tietz N W et al. Clinical Guide to Laboratory Tests, 3rd ed. AACC 1995.

